



Certificate of Analysis

Document No.: MB-BRP-ZL-03

Version No.: A/0

Mycobacteria DNA Detection Kit

Product No.: **1503602**

Lot No.: **20240501**

Expiration Date: **31 Oct. 2025**

Quality Control Data:

Test Item	Specification	Result
Low concentration sample	Results of 40 copies/reaction sample: Ct value of FAM channel < 40.00, and Ct value of VIC channel < 35.00.	PASS
Positive Control	Ct value of FAM channel < 35.00, and Ct value of VIC channel < 35.00.	PASS
No Template Control (NTC)	Ct value of FAM channel \geq 40.00 or absence of specific amplification, and the Ct value of VIC channel < 35.00.	PASS

Quality Assurance Wang Bingwang

DAY/MONTH/YEAR 28 May 2024
Huzhou Shenke Biotechnology Co., Ltd.

Certificates available online, please access www.shenkebio.com/en

For any other inquiries, please contact us by email: info@shenkebio.com

Address: 8th Floor, 6B Building, No. 1366 Hongfeng Road, Huzhou 313000, Zhejiang Province, China