



Certificate of Analysis

Document No.: MB-BRP-ZL-10

Version No.: A/0

Mycobacteria DNA Extraction Kit

Product No.: **1503601**

Lot No.: **20240501**

Expiration Date: **31 Oct. 2024**

Quality Control Data:

Test Item	Specification	Result
Positive Sample (with matrix)	Mycobacteria DNA in test sample (with matrix) was extracted and analyzed. Ct value of FAM channel < 40.00, and Ct value of VIC channel < 35.00.	PASS
Positive Sample (without matrix)	Mycobacteria DNA in test sample (without matrix) was extracted and analyzed. Ct value of FAM channel < 40.00, and Ct value of VIC channel < 35.00.	PASS
Negative Control Sample	Ct value of FAM channel \geq 40.00 or absence of specific amplification, and Ct value of VIC channel < 35.00.	PASS

Quality Assurance Wang Qingwang

DAY/MONTH/YEAR 27 May 2024
Huzhou Shenke Biotechnology Co., Ltd.

Certificates available online, please access www.shenkebio.com/en

For any other inquiries, please contact us by email: info@shenkebio.com

Address: 8th Floor, 6B Building, No. 1366 Hongfeng Road, Huzhou 313000, Zhejiang Province, China