



Certificate of Analysis

Document No.: Mycop-BRP-ZL-26

Version No.: A/0

Mycoplasma DNA Detection Kit (2G)

Product No.: **1509841**

Lot No.: **20240502**

Expiration Date: **20 Dec. 2025**

Quality Control Data:

Test Item	Specification	Result
Positive Control	Ct value of FAM channel < 35.00, and Ct value of VIC channel < 35.00.	PASS
No Template Control (NTC)	Results of NTC: Ct value of FAM channel ≥ 40.00 or absence of specific amplification, and the Ct value of VIC channel < 35.00.	PASS

Quality Assurance Wang Qingwang

DAY/MONTH/YEAR 28 May 2024
Huzhou Shenke Biotechnology Co., Ltd.

Certificates available online, please access www.shenkebio.com/en

For any other inquiries, please contact us by email: info@shenkebio.com

Address: 8th Floor, 6B Building, No. 1366 Hongfeng Road, Huzhou 313000, Zhejiang Province, China