



Certificate of Analysis

Document No.: Mycop-BRP-ZL-29

Version No.: A/0

Mycoplasma DNA Extraction Kit (2G)

Product No.: **1509840**

Lot No.: **20240502**

Expiration Date: **05 Mar. 2025**

Quality Control Data:

Test Item	Specification	Result
Positive Sample (with matrix)	Mycoplasma DNA in test sample (with matrix) was extracted and analyzed. Ct value of FAM channel<40.00, and Ct value of VIC channel<35.00.	PASS
Positive Sample (without matrix)	Mycoplasma DNA in test sample (without matrix) was extracted and analyzed. Ct value of FAM channel<40.00, and Ct value of VIC channel<35.00.	PASS
Negative Control Sample	Mycoplasma DNA of Negative control sample was extracted and analyzed. Ct value of FAM channel≥40.00 or absence of specific amplification, and Ct value of VIC channel<35.00.	PASS

Quality Assurance

Wang Qinyang

DAY/MONTH/YEAR

28 May 2024

Huzhou Shenke Biotechnology Co., Ltd.

Certificates available online, please access www.shenkebio.com/en

For any other inquiries, please contact us by email: info@shenkebio.com

Address: 8th Floor, 6B Building, No. 1366 Hongfeng Road, Huzhou 313000, Zhejiang Province, China