



# Certificate of Analysis

Document No.: CHO-BRP-ZL-14

Version No.: A/0

## Residual CHO DNA Quantitation Kit (2G)

Product No.: 1101100-1

Lot No.: 20240501

Expiration Date: 26 Oct. 2025

### Quality Control Data:

Test Item	Specification	Result
Linearity of Standard curve	$R^2 \geq 0.990$ .	PASS
Amplification efficiency of Standard curve	$E = 90.0\% - 110.0\%$ .	PASS
Repeatability	3 fg/ $\mu$ L CHO DNA Control was detected, $CV \leq 15.0\%$ .	PASS
No Template Control (NTC)	The Ct value of NTC should be higher than 35.00 or absence of specific amplification.	PASS

Quality Assurance Wang Qingwang

DAY/MONTH/YEAR 28 May 2024  
Huzhou Shenke Biotechnology Co., Ltd.

Certificates available online, please access [www.shenkebio.com/en](http://www.shenkebio.com/en)

For any other inquiries, please contact us by email: [info@shenkebio.com](mailto:info@shenkebio.com)

Address: 8th Floor, 6B Building, No. 1366 Hongfeng Road, Huzhou 313000, Zhejiang Province, China