



# Certificate of Analysis

Document No.:E.coli RNA-BRP-ZL-08

Version No.:A/0

## Residual *E.coli* RNA Quantitation Kit(2G)

Product No.: 1201201-1

Lot No.: 20240501

Expiration Date: 09 Oct. 2025

### Quality Control Data:

Test Item	Specification	Result
Linearity of Standard curve	$R^2 \geq 0.990$ .	PASS
Amplification efficiency of Standard curve	$E = 83.3\% - 110.0\%$ .	PASS
Repeatability	2 fg/ $\mu$ L <i>E.coli</i> RNA Control was detected, $CV \leq 15.0\%$ .	PASS
No Template Control (NTC)	The average Ct value of NTC should be at least 2 cycles higher than the lowest standard curve.	PASS

Quality Assurance Wang Qingwang

YEAR/MONTH/DAY 28 May 2024  
Huzhou Shenke Biotechnology Co., Ltd.

Certificates available online, please access [www.shenkebio.com/en](http://www.shenkebio.com/en)

For any other inquiries, please contact us by email: [info@shenkebio.com](mailto:info@shenkebio.com)

Address:8th Floor,6B Building, No. 1366 Hongfeng Road, Huzhou 313000, Zhejiang Province, China