



Certificate of Analysis

Document No.: HEK-BRP-ZL-07

Version No.: A/0

Residual HEK293 DNA Quantitation Kit (3G)

Product No.: 1101104-1

Lot No.: 20240501

Expiration Date: 17 Apr. 2025

Quality Control Data:

Test Item	Specification	Result
Linearity of Standard curve	$R^2 \geq 0.990$.	PASS
Amplification efficiency of Standard curve	$E = 90.0\% - 110.0\%$.	PASS
Repeatability	30 fg/ μ L HEK293 DNA Control was detected, $CV \leq 15.0\%$.	PASS
No Template Control (NTC)	The mean value of NTC should not exceed 6.00 fg/ μ L.	PASS

Quality Assurance

Wang Qingwang

DAY/MONTH/YEAR

28 May 2024

Huzhou Shenke Biotechnology Co., Ltd.

Certificates available online, please access www.shenkebio.com/en

For any other inquiries, please contact us by email: info@shenkebio.com

Address: 8th Floor, 6B Building, No. 1366 Hongfeng Road, Huzhou 313000, Zhejiang Province, China