



# Certificate of Analysis

Document No.: DNA-BRP-ZL-24

Version No.: A/0

## Residual Host Cell DNA Sample Preparation Kit

Product No.: 1104191

Lot No.: 20240501

Expiration Date: 28 Feb. 2025

### Quality Control Data:

Test Item	Specification	Result
Recovery rate	The recovery rate of samples was between 70.0% -130.0%.	PASS
Extracted Positive Control	The recovery rate of Positive control was between 80.0%-120.0%.	PASS
Negative Control Sample (NCS)	The Ct value of NCS should be higher than the mean Ct value of the lowest concentration in the standard curve.	PASS

Quality Assurance Wang Qingwang

DAY/MONTH/YEAR 28 May 2024  
Huzhou Shenke Biotechnology Co., Ltd.

Certificates available online, please access [www.shenkebio.com/en](http://www.shenkebio.com/en)

For any other inquiries, please contact us by email: [info@shenkebio.com](mailto:info@shenkebio.com)

Address: 8th Floor, 6B Building, No. 1366 Hongfeng Road, Huzhou 313000, Zhejiang Province, China