



Certificate of Analysis

Document No.: E.coli RNA-BRP-ZL-11

Version No.: A/0

Residual Host Cell RNA Sample Preparation Kit

Product No.: 1201205

Lot No.: 20240501

Expiration Date: 02 Feb. 2025

Quality Control Data:

Test Item	Specification	Result
Linearity of Standard curve	$R^2 \geq 0.990$.	PASS
Amplification efficiency of Standard curve	$E = 83.3\% - 110.0\%$.	PASS
Recovery rate	The recovery rate of samples was between 50.0%-150.0%.	PASS
Negative Control Sample (NCS)	The average Ct value of NCS should be at least 2 cycles higher than the lowest standard curve.	PASS
No Template Control (NTC)	The average Ct value of NTC should be at least 2 cycles higher than the lowest standard curve.	PASS

Quality Assurance Wang Bingwang

DAY/MONTH/YEAR 28 May 2024
Huzhou Shenke Biotechnology Co., Ltd.

Certificates available online, please access www.shenkebio.com/en

For any other inquiries, please contact us by email: info@shenkebio.com

Address: 8th Floor, 6B Building, No. 1366 Hongfeng Road, Huzhou 313000, Zhejiang Province, China