



Certificate of Analysis

Document No.: Vero-BRP-ZL-10

Version No.: A/0

Residual Vero DNA Size Analysis Kit (2G)

Product No.: 1103174

Lot No.: 20240501

Expiration Date: 09 Mar. 2025

Quality Control Data:

| Test Item | Specification | Result | |
|--|--|--|------|
| Linearity of Standard curve | $R^2 \geq 0.990$. | PASS | |
| Amplification efficiency of Standard curve | $E = 83.3\% - 110.0\%$. | PASS | |
| Repeatability | Vero Primer&Probe MIX-85 | 30 fg/ μ L Vero DNA Control was detected, $CV \leq 20.0\%$. | PASS |
| | Vero Primer&Probe MIX-134 | | PASS |
| | Vero Primer&Probe MIX-229 | | PASS |
| | Vero Primer&Probe MIX-552 | | PASS |
| No Template Control (NTC) | The Ct value of NTC should be 2 larger than the Ct value of ST5. | PASS | |

Quality Assurance Wang Qingwang

DAY/MONTH/YEAR 28 May 2024
Huzhou Shenke Biotechnology Co., Ltd.

Certificates available online, please access www.shenkebio.com/en

For any other inquiries, please contact us by email: info@shenkebio.com

Address: 8th Floor, 6B Building, No. 1366 Hongfeng Road, Huzhou 313000, Zhejiang Province, China